

**Intox EC/IR-II
Quick Test**

RANDOM DRUG UNIT
Production

Serial Number: [REDACTED]

Test Number: 1142

Test Date: 12/16/2011

Test Time: 10:44 CST

Comments: [REDACTED]

System Check: Passed

Test	g/210L	Time
BLK	.000	10:46
SUBJ	.000	10:47

Test Status: Success

=====

**Intox EC/IR-II
Scheduled
Certification**

RANDOM DRUG UNIT
Production

Serial Number: [REDACTED]

Test Number: [REDACTED]

Test Date: 12/01/2011

Test Time: 07:00 CST

Dry Gas Target: .080

Lot Number: [REDACTED]

Exp Date: 06/28/2012

System Check: Passed

Test	g/210L	Time
BLK	.000	07:01
CHK	.080	07:01
BLK	.000	07:02
CHK	.080	07:03

Test Status: Success